





检测报告

(Test Report)

No. BOYGFLAR873347L1

样品名称 (Sample Description) 一次性使用医用口罩 Disposable Medical Face Mask

委托单位 (Applicant) 浙江紫家防护用品有限公司 Zhejiang The Purples Protective Products Co.,Ltd.





- 1. 本报告无专用章和批准人签章无效。
 - This report is invalid without the approver's signatures and special seal of inspection.
- 2. 本报告页面所使用 "PONY"、"谱尼"字样为本单位的注册商标,其受《中华人民共和国商标法》保护,任何未经本单位授权的擅自使用和仿冒、伪造、变造"PONY"、"谱尼"商标均为违法侵权行为,本单位将依法追究其法律责任。The pattern and characters of "PONY" and "谱尼" used in this report are protected by the trademark law of the People's Republic of China. Any unauthorized usage, counterfeit, forgery and alteration of trademarks of "PONY" and "谱尼" are the violations of the law. The PONY has the right to pursueall legal liabilities of the subject of the delict.
- 委托单位对报告数据如有异议,请于报告完成之日起十五日内(初级农产品报告请于报告收到之日起五日内)向本单位书面提出复测申请,同时附上报告原件并预付复测费。

If the applicant has any questions about the results, shall provide a written retest application with the original report, and prepay the retest fees to PONY within fifteen days since the approval date (as an exception, it shall be within five days since the date received for the primary agriculture products report).

- 4. 委托单位办理完毕以上手续后,本单位会尽快安排复测。如果复测结果与异议内容相符,本单位将退还委托单位的复测费。 After the applicant finishes the procedure mentioned above, PONY shall arrange the retest as soon as possible. If the retest result accords with the applicant dissent, PONY shall refund the retest fees.
- 5. 不可重复性或不能进行复测的实验,不进行复测,委托单位放弃异议权利。 Tests that can not be repeated and tested shall not be carried out again.
- 6. 委托单位对样品的代表性和资料的真实性负责,否则本单位不承担任何相关责任。 The applicant should undertake the responsibility for the provided samples' representativeness and document authenticity. Otherwise, PONY has not any relevant responsibilities.
- 本报告仅对所测样品负责,报告数据仅反映对所测样品的评价,对于报告及所载内容的使用、使用所产生的直接或间接损失及一切法律后果,本单位不承担任何经济和法律责任。

This report is only responsible for the provided sample. The test results only represent the evaluation of the tested sample. PONY will not be responsible for any economical or legal liability generated from direct or indirect usage of the test report.

- 8. 本单位有权在完成报告后处理所测样品。
 - PONY has the right to dispose the tested sample after approval of the test report.
- 9. 本单位保证工作的客观公正性, 对委托单位的商业信息、技术文件等商业秘密履行保密义务。 PONY assures objectivity and impartiality of the test, and fulfills the obligation of confidentiality for applicant's commercial information, and technique document.
- 10. 本报告私自转让、盗用、冒用、涂改、未经本单位批准的复制(全文复制除外)或以其它任何形式的篡改均属无效、本单位 将对上述行为严究其相应的法律责任。

The report is invalid in case of illegal transfer, embezzlement, imposture, modification or any altering, reproducing except in full, without approval of PONY. PONY shall investigate and affix the applicant's legal liability accordingly.

▲ 防伪说明 (Anti-counterfeiting Description):

- (1) 报告编号是唯一的;
 - The test report has exclusive report code.
- (2) 报告采用特制防伪纸张印制,纸张表面带有"PONY"防伪纹路,该防伪纹路不支持复印,即复制件不会带有"PONY" 防伪纹路。

The test report is printed by anti-copying paper whose surface shows "PONY" security print with specific anticounterfeiting technique. Security print will disappear after copying. Duplicates are not expected to give "PONY" security print under any



全国服务热线 400-819-5688

WWW.PONYTEST.COM

扫描二维码 歐彩唱 关注谱尼测试微信公众号



北京实验室: (010) 83055000 武汉实验室: (027) 83997127 哈尔滨实验室: (0451)58627755 上海实验室: (021) 64851999 长春实验室: (0431)85150908 石家庄实验室: (0311)85376660 温州实验室: (0577)88271060 合肥实验室: (0551)63843474 青岛实验室: (0532)88706866 大连实验室: (0411)87336618 乌鲁木齐实验室: (0991) 6684186 深圳实验室: (0755)26050909 郑州实验室: (0371)69350670 呼和浩特实验室: (0471) 3450025 广州实验室: (020) 89224310

天津实验室: (022) 23607888 西安实验室: (029) 89608785 杭 州 实 验 室 : (0571)87219096 苏州实验室: (0512)62997900 太原实验室: (0351) 7555762 宁 波 实 验 室: (0574)87736499

厦门实验室: (0592)5568048 成都实验室: (028) 87702708



检测结果

(Test Results)

No. BOYGFLAR873347L1

第1页, 共2页 (page 1 of 2)

一次性使用医用口罩	17 8 18 15	<	把			
Disposable Medical Face Mask	样品规格 (Sample Specification)		规格 Specification: 17.5*9.5cm 型号 Model: TP-100			
浙江紫家防护用品有限 公司 Zhejiang The Purples Protective Products Co.,Ltd.	商标 (Trade Mark)					
2020-04-09	生产日期或批号		生产日期 Manufacturing Date: 2020-04-02 生产批号 Manufacturing Lot No.: 20200402Z			
2020-04-09~2020-04-23	样品等级 (Sample Grade)					
150pcs	检测类别 (Test Type)		委托检测 (Commissioning Test)			
正常 Normal	检测环境 (Test Environment)		符合要求 (To meet the requirements)			
见下页 See the next page						
见下页 See the next page						
电热恒温培养箱 Electric Heating Constant Temperature Incubator、 细菌过滤效率检测仪 Bacterial filtration efficiency detector 等 etc.						
限值标准 Limit Standard: EN 14683: 2019 生产单位 Production: 浙江紫家防护用品有限公司 Zhejiang The Purples Protective Products Co.,Ltd. 受检单位 Inspection: 浙江紫家防护用品有限公司 Zhejiang The Purples Protective Products Co.,Ltd. 以上样品信息由委托单位提供 The information of above sample was provided by the applicant.						
编制人 (Edited by)		1分分				
审核人 (Checked by)		郭惠				
批准人 (Approved by)		戴隆				
签发日期 (Issued Date)		2020-04-23				
	浙江紫家防护用品有限公司 Zhejiang The Purples Protective Products Co.,Ltd. 2020-04-09 2020-04-09 2020-04-09~2020-04-23 150pcs 正常 Normal 见下页 See the next page 见下页 See the next page 电热恒温培养箱 Electric F细菌过滤效率检测仪 Back 限值标准 Limit Standard: 生产单位 Production: 浙Protective Products Co.,Ltd 受检单位 Inspection: 浙Protective Products Co.,Ltd 以上样品信息由委托单位 The information of above s编制人 (Edited by) 审核人 (Checked by) 批准人 (Approved by)	新江紫家防护用品有限公司 Zhejiang The Purples Protective Products Co.,Ltd. 2020-04-09 (Manufacturing Da Lot No.) 2020-04-09~2020-04-23	勝江紫家防护用品有限 Zhejiang The Purples Protective Products Co.,Ltd. 2020-04-09 (Manufacturing Date or Lot No.) (Sample Grade) (Sample Grade) (Test Type) 正常 Normal ル下页 See the next page 現下页 See the next page 中熱恒温培养箱 Electric Heating Constant Temperatu 细菌过滤效率检测仪 Bacterial filtration efficiency del 限值标准 Limit Standard: EN 14683; 2019 生产单位 Production: 浙江紫家防护用品有限公司Protective Products Co.,Ltd. 受检单位 Inspection: 浙江紫家防护用品有限公司Protective Products Co.,Ltd. 以上样品信息由委托单位提供 The information of above sample was provided by the 编制人 (Edited by) 事核人 (Checked by) 批准人 (Approved by) 签发日期			

© Hotline 400-819-5688 www.ponytest.com 谱尼测试集团上海有限公司

公司地址: 上海市松江区文翔东路 99 号 7 幢 2 层

电话: 021-37895599

测试地址:上海市松江区文翔东路99号5幢、6幢、7幢1层、7幢3层

上海市徐汇区桂平路 680 号 35 幢 2-4 楼、6 楼



检测结果

(Test Results)

No. BOYGFLAR873347L1

第2页, 共2页 (page 2 of 2)

样品名称和编号 (Sample Description and Number)	检测项目 (Test Items)	限值 (Limit)		检测结果		检测方法	
		I 型 Type I	II型 Type II	II R 型 Type II R	(Test Results)		(Test Methods)
一次性使用医用 efficiency(BFE)	(BFE), % Bacterial filtration	≥95	≥98	≥98	No.1	99.0	EN 14683: 2019
					No.2	98.9	
					No.3	99.0	
					No.4	99.0	
					No.5	98.9	
	5 TH 100 CH 100	<40	<40	<60	29.9		EN 14683: 2019
	cfu/g Microbial	€30	€30	€30	1		EN 14683: 2019

样品编号和照片 (Sample Number and Photo):



仅对报告照片中的样品负责 Pony authenticate the photo on original report only —以下空白— (End of Report)

